

TOPICAL REVIEW • OPEN ACCESS

Time-gated Raman spectroscopy – a review

To cite this article: Martin Kögler and Bryan Heilala 2020 *Meas. Sci. Technol.* **32** 012002

View the [article online](#) for updates and enhancements.

You may also like

- [Wide-field time-gated SPAD imager for phasor-based FLIM applications](#)
Arin Ulku, Andrei Ardelean, Michel Antolovic et al.
- [Impact of silicide layer on single photon avalanche diodes in a 130 nm CMOS process](#)
Zeng Cheng, Darek Palubiak, Xiaoqing Zheng et al.
- [High-throughput, multi-parametric, and correlative fluorescence lifetime imaging](#)
Chetan Poudel, Ioanna Mela and Clemens F Kaminski

Topical Review

Time-gated Raman spectroscopy – a review

Martin Kögler^{1,3}  and Bryan Heilala² ¹ VTT Technical Research Centre of Finland, Oulu, Finland² Timegate Instruments Oy, Oulu, FinlandE-mail: martin.kogler@vtt.fi

Received 7 May 2020, revised 16 July 2020

Accepted for publication 18 August 2020

Published 27 October 2020



CrossMark

Abstract

Time-gated (TG) Raman spectroscopy (RS) has been shown to be an effective technical solution for the major problem whereby sample-induced fluorescence masks the Raman signal during spectral detection. Technical methods of fluorescence rejection have come a long way since the early implementations of large and expensive laboratory equipment, such as the optical Kerr gate. Today, more affordable small sized options are available. These improvements are largely due to advances in the production of spectroscopic and electronic components, leading to the reduction of device complexity and costs. An integral part of TG Raman spectroscopy is the temporally precise synchronization (picosecond range) between the pulsed laser excitation source and the sensitive and fast detector. The detector is able to collect the Raman signal during the short laser pulses, while fluorescence emission, which has a longer delay, is rejected during the detector dead-time. TG Raman is also resistant against ambient light as well as thermal emissions, due to its short measurement duty cycle.

In recent years, the focus in the study of ultra-sensitive and fast detectors has been on gated and intensified charge coupled devices (ICCDs), or on CMOS single-photon avalanche diode (SPAD) arrays, which are also suitable for performing TG RS. SPAD arrays have the advantage of being even more sensitive, with better temporal resolution compared to gated CCDs, and without the requirement for excessive detector cooling. This review aims to provide an overview of TG Raman from early to recent developments, its applications and extensions.

Keywords: time-gated, time-resolved, Raman spectroscopy, SPAD detector, pulsed laser excitation, surface-enhanced Raman spectroscopy

(Some figures may appear in colour only in the online journal)

1. Introduction

The Achilles' heel of conventional Raman spectroscopy (RS) is sample-induced fluorescence emission. This is a

competing phenomenon, occurring with relatively weak Raman scattering, and can obscure the whole Raman spectrum, making the identification or quantification of materials impossible. An effective solution to this problem is time-gating (TG), which is a general technique used in signal processing [1]. The purpose of TG spectroscopy is the measurement of signals during a specific portion of time, which enables the monitoring of transient processes [2]. TG found its way into RS as early as in the 1970s, as scientists sought out ways to remove fluorescence background signals during measurements, cf. chapter 1.3 [3]. However, TG Raman

³ Author to whom any correspondence should be addressed.



Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](https://creativecommons.org/licenses/by/4.0/). Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.

has only become commercially available in recent years. To extend the applicability of RS in general, it is important to overcome the fluorescence limitation.

RS is based on the inelastic scattering of photons shifted from the excitation wavelength, known as Stokes and Anti-Stokes shifts [4]. It is employed to provide information about excited molecules in a given sample. Similarly to infrared spectroscopy (IR), this information can be used to study the chemical or biological fingerprint of materials at different states of aggregation (solid, liquid, or gas) [5, 6]. However, band intensities and selection rules are important differences between the two vibrational spectroscopic techniques. In RS, the transition of polarizability of the molecules is shifted from the excitation wavelength, whereas IR relates to the transitional dipole moment [7]. Commonly a monochromatic excitation light source (laser) is used for RS, whereas for IR, a broader excitation source (LED or halogen lamp) can be used. The fundamental advantage of RS over IR is that it can be used to investigate liquid or moist samples without strong interference from water responses. The two techniques often complement one another, if the concentration of water in the sample is low. Overall, the applicability of any analytical technique is also dependent on the properties of the sample itself, as the optical scattering efficiencies vary in solid materials, particles in liquids, and droplets/bubbles in liquids, which can, for example, cause photons to be scattered multiple times, making quantitative and qualitative analyses challenging [8].

In general, both vibrational spectroscopic methods allow measurement results to be rapidly obtained, unlike many other analytical methods [8]. However, RS provides the full set of rotational and vibrational spectral information, which is otherwise only available by combining mid-infrared and far-infrared detection techniques [9]. Unfortunately, Raman scattering is weak: from about 10^7 laser excited photons, only one Raman photon is inelastically scattered [7]. Sufficient RS signal intensities can be achieved by increasing laser power to a moderate level, optimizing the detector integration time, and ensuring that laser and collection spots are properly focused to the target while not damaging the sample. The quality of the detected Raman signal depends on several factors, relating either to the instrument itself, the sample under inspection, the setup, or the surrounding conditions. These factors can include (a) the detector itself, or detector shot noise that can be limited when the spectrometer detector is cooled and/or temperature stabilized, (b) the sporadic but high intensity detection of cosmic rays, (c) black-body radiation (d) ambient light such as LEDs, or incandescent bulbs, as well as mercury-vapor or gas-discharge lamps, and (e) fluorescence and other types of photoluminescence interference [10–14]. A common solution for performing Raman measurements without ambient light interference is to measure in a dark space, or by placing the sample in a stray light sealed sample enclosure. The broad spectral disturbance of fluorescence in Raman measurements represents a major challenge across all fields where RS is currently employed, and limits its wider application [7]. For example, the low Raman effective cross-section per molecule (Raman scattering of around 10^{-31} to 10^{-29} cm²) is dependent on λ_{exc} (excitation wavelength); where the surrounding refraction

index (sample medium) to fluorescence effective cross-section per molecule is approximately 10^{-16} cm², it is clearly difficult to obtain viable Raman measurement results with strongly fluorescent samples [5, 15].

Fluorescence background may originate from impurities in the sample/solvent, the matrix components of samples (especially where these are coloured), or the analyte itself [14]. Fluorescence background may also originate from optical components in the pathway of the spectrometer, such as lens coatings. Occasionally, fingerprints on lenses or on fibre connectors may cause fluorescence. The Raman signal strength (I) is a function of the fundamental laser wavelength. It is proportional to the fourth power of the laser wavelength (λ^4), the intensity of the laser radiation (I_L), the number density of scattering molecules (N) — of which a major part may originate from sample-induced fluorescence, cf. equation (2) — and the polarizability change (δ_α/δ_q) as described in equation (1):

$$I \sim f^4 \times I_L \times N \times \left(\frac{\delta_\alpha}{\delta_q} \right) \sim \left(\frac{1}{\lambda^4} \right) \quad (1)$$

[10, 16]. The strength of the Raman signal increases with the fourth power of the frequency of the incident light. Raman and fluorescence emissions are high in the visible spectral region (around 530 nm), as shown in figure 1, but decrease towards the near infrared spectral range [7, 10].

Figure 1 can also be extended to include shorter wavelengths i.e. the UV spectral range, where at very short wavelengths, fluorescence is no longer a problem, but the risk of sample degradation generated by the UV laser is increased [7]. A common practical compromise between the amount of available Raman scattering and fluorescence reduction is the use of a 785 nm laser excitation wavelength with corresponding Raman spectrometer setup. However, this setup may not be applicable for high fluorescent samples, as discussed below, in the *application and extension* sections of this TG Raman review. According to the Perrin-Jablonski molecular energy level diagram, the fluorescence process itself is determined by excitation, conversion and emission, which occur at different times. There are three stages operate as follows: (i) the excitation of fluorophore molecules by re-radiating photons, which occurs within femtoseconds, (ii) in approximately the same time frame, non-radiative internal conversion processes take place due to vibrational relaxation, and (iii) detectable fluorescence emission occurs over a much slower time frame, i.e. around the pico- to nano-second scale, depending on the sample [17].

The main aim of the TG principle in RS is to suppress sample-induced fluorescence, as well as phosphorescence, during the measurement process, and to keep the signal-to-noise ratio (SNR) sufficiently high while simultaneously suppressing other potential continuous disturbances, such as ambient light and thermal emissions [18].

$$\text{SNR} = \frac{N_{\text{Raman}}}{\sqrt{(N_{\text{Raman}} + N_{\text{Fluorescence}} + N_{\text{Dark count rate}})}} \quad (2)$$




Figure 1. Raman signal strength and influence of the fourth-power rule (*) with respect to common Raman excitation wavelengths and the impact of fluorescence emission, modified with permission from ref [10].

The SNR, as shown in equation (2), can be maximized by tuning the width and position of the time-gate, whereas (N) is the number density of scattering molecules for the Raman, fluorescence, and the detector's dark count rate, respectively (cf. equation (1) and equation (2) [10, 19]). The relation of Raman versus fluorescence photons in terms of the SNR seems obvious. While Raman scattering has a very short lifetime, the fluorescence process involves real electronic excited states with finite measurable lifetimes; thus the Raman and fluorescence signal can be separated in the temporal domain [7]. This review highlights key findings and applications relating to past, present, and future developments of TGRS.

1.1. TG Raman spectrometer setup and measurement principle

The components and typical setup of a TG Raman spectrometer are illustrated in figure 2(a). The essential component for performing TG Raman is a pulsed laser excitation source with a suitable repetition rate, pulse width and pulse energy. Most of the pulsed laser energy is focussed onto the sample spot for the purpose of excitation, but a fraction is used to enable TG through the delay generator, matching the sequence of detection, and used for temporal synchronization with the detector. The main components are as follows: a pulsed laser (commonly in picosecond time-frame) with a fast repetition rate (typically in MHz range), and a delay generator, synchronized to the detector-spectrograph unit via opto-electrical adjustable delay settings, and to a computer, which acts as a controller and measuring device. Figure 2(b) shows the temporal profile of TGRS, with its adjustable time-gate and concomitant rejection of fluorescence [10, 17, 20–22]. Based on the operational principle depicted in figure 2(a), the detector is activated only during the emission pulses, as shown in figure 2(b). The difference between TG (dashed line) and CW (continuous line) operation mode, with a fictive Raman spectrum for each mode, is illustrated in figure 2(c).

Until recently, the complexity, size, and price of TG Raman spectrometers have precluded its widespread use. Different types of detectors have been employed in the past, such as multichannel plates (MCPs), photo-multiplier tubes (PMTs), and ultra-fast and intensified charge-coupled devices (ICCDs), with and w/o Kerr gates. Essentially, these devices require detector cooling, and are thus very complex and bulky, (see table 1) [24]. Since the turn of the millennium, complementary-metal-oxide-semiconductor (CMOS) single-photon avalanche diode (SPAD) array detectors have been commercially available [21, 24]. CMOS SPADs have the benefit of significantly reducing the disadvantages mentioned above. Moreover, there is often no need to cool the detector, which further reduces the complexity, cost, and overall size of the spectrometer. Current commercial TG Raman devices are portable tabletop sized units, suitable for process industry monitoring purposes [25]. We compare different detector technologies suitable for TG in section 3, (see table 1).

The terms *time-resolved* (TR) and TG are often used interchangeably. However, while TG mainly focusses on the suppression of fluorescence and other disturbances, thereby repeating a cycle of measurements across the width or part of the width of the pulsed excitation source, it can also be used to perform TR measurements in terms of studying transient processes. Due to its continuous temporal repetition, a third measurement dimension (time) is created, as shown in figure 3. The fixed temporal delay is very often used to study transient change and fluorescence decay processes. Both TG and TR have similar technical features which measure specified temporal regions, rejecting unwanted emissions, and thus enabling the separation of Raman and fluorescence [27]. Although in evaluating the literature for this review, it seems that the majority of articles describing the technical aspects of fluorescence signal suppression are actually referring to TG rather than TR, we feel that it would be inappropriate to view these terms as interchangeable.




Figure 2. Time-gated Raman spectroscopy—(a) general block-diagram of a TG-spectrometer (b) Temporal profile with time-gating, detection of Raman photons (red-dotted line), rejection of most fluorescence photons (green-dotted line) with residual photons (blue-dotted line) and (c) TG Raman (green-dotted line) compared to conventional—continuous wave Raman (green) in spectral domain for the case of a strong fluorescent sample measurement, (b) and (c) are modified and reprinted with permission from refs. [10, 23].




Figure 3. Sample TG Raman data: (a) Raman signal of sesame oil as a function of time (sampling shown as 3D-data-cube) with fluorescence interference, and (b) sample-specific fluorescence lifetime, reprinted with permission from reference [10].

1.2. TG Raman measurement result

A TG Raman measurement result is a three-dimensional plot of intensity versus frequency and time; - it is therefore possible to view Raman and fluorescence simultaneously [28]. The raw data temporally sampled after each laser pulse is subsequently accumulated, as illustrated in figure 3(a).

Sesame oil, shown in figure 3(a), exhibits a high fluorescence background and a lifetime of more than 4 ns. It can be used to demonstrate the effectiveness of TG Raman, as compared to conventional CW Raman [19]. Figure 3(b) shows several fluorescence characteristics and lifetimes for various samples [29]. It also demonstrates that setting a corresponding gate width is important, in that each sample has

Table 1. Summary of prominent fluorescence suppression methods, modified and extended from references. [10, 17, 26].

Method	Time-domain	Frequency-domain	Wavelength-domain	Computational
Cost	Medium to high	Medium to high	Medium	Low
Performance	High	Medium	Medium	Medium to high
Advantages	Separation of Raman signal from fluorescence and other background interference. Minimal risk of photon scattering effect in thick samples. Effective statistical noise reduction. Short duty cycle for excitation and detection system. A third dimension enables the representation of data in the temporal domain.	Combination of fast responding detectors, e.g. PMTs and spectrum analyzers is possible, which can distinguish high frequencies, primarily from Raman (coupled to excitation source), and lower frequencies, primarily from fluorescence signals. Fluorescence can be reduced by phase-nulling or demodulation.	Require only moderate modification of conventional Raman systems.	After subtracting the background (i.e. background correction), specific spectral features can be preserved.
Disadvantages	Ineffective separation between Raman and fluorescence when fluorescence lifetime is comparable to the excitation pulse duration (timing window). Spectral resolution is affected by the laser pulse width. The Kerrgate: requires very high pulse energies to operate.	High intensity Raman emission required to overcome fluorescence interference. Any noise on the fluorescence signal (e.g. shot noise) within the detector bandwidth will contribute to the phase-nulled Raman spectrum.	Complex data processing is required sequentially after the performance of different Raman measurements, to obtain the normal Raman spectra. Strong dependence of spectral reproduction on the noise level of the raw spectra.	Optimal choices of key parameters are user-experience dependent. Time consuming data processing, which in some circumstances results in distorted spectral features that have little in common with the original spectral information. Dominance of artifacts, especially in weak Raman signals. Sharp Raman peaks can be left out due to random variations in pixel-to-pixel sensitivity.

different sample-induced fluorescence emission characteristics.

1.3. Brief historical overview of the development of TG Raman

TG Raman has its roots in the 1970s, with the advent of pulsed laser systems in the subnanosecond range, and the availability of computer controllable detectors. In 1972, Perry P. Yaney was the first to experimentally and theoretically prove the concept of fluorescence interference reduction by applying the TG principle [30]. Another pioneer was Richard Van Duyne, who is also one of the discoverers of the surface-enhanced Raman effect (see Chapter 5 of this review [3]). In 1974, the Van Duyne research group demonstrated experimentally for the first time that fluorescence rejection was possible, using a Rhodamine 6 G dye-doped benzene sample, while improving the SNR with the combination of a photomultiplier tube (PMT) and a pulsed argon ion laser source at 488 nm excitation [3]. In 1976, Yaney, using a similar setup to Van Duyne *et al.*, but with a different pulsed excitation source (ps pulsed Nd:YAG, neodymium-doped yttrium aluminium garnet laser at 532 nm), found that TG Raman significantly improved the spectral results of three major Raman bands of acridine orange in benzene with a short laser pulse width (around 200 ns), as compared to CW Raman. He also stated that ambient light does not interfere with TG Raman spectral results, and that it improved the SNR of weak Raman signals in the presence of

fluorescence. In addition, he noted that simultaneous fluorescence processes in a sample limited Raman detection, and that the TG principle enabled the use of short gating times, and higher dark current detectors are acceptable, e.g. uncooled PMTs [18]. In the same year, Harries *et al* were the first team to compare the level of fluorescence background rejection in a TR experiment to that of continuous excitation on a fluorophore-doped benzene Raman band at 992 cm^{-1} [31].

Laser systems and detectors at this time required large, complex devices, requiring very precise device calibrations [2]. By 1985, Deffontaine *et al* were testing active and passive methods for picosecond (ps) time gating, with the aim of improving SNR, in combination with synchronous streak camera detection and an optical Kerrgate; however, they noted the limited applicability of this approach [32]. In the same year, Watanabe *et al* demonstrated fluorescence rejection with ethanol doped Rhodamine 6 G in an ultra-short TG window of 31 ps, with a fast gating PMT-MCP arrangement and a 570 nm ps pulsed laser [33]. One year later, in 1986, both Overall *et al* and Howard *et al* demonstrated independently that with ps pulsed lasers systems, the bulk of fluorescence photons from a dye (rubrene) could be rejected from the pure Raman signal of benzene when the detection of the MCP was performed in TG mode [22, 34]. The ps pulsed laser excitation and MCP acquisition was limited with the 8-bit computers of the time, and therefore only one Raman band, i.e. benzene (992 cm^{-1}), was observed, while rubrene was sequentially added. Prior to this, fluorescence-rejection had only been

tested on non-fluorescent samples, such as benzene or hexane, which were doped with fluorescent dyes. Only the most prominent Raman bands of these samples could be investigated, rather than the whole spectrum. Various applications for TG Raman emerged in the following years. For example Petrich *et al* examined the resonance TR Raman of haemoglobins with an Nd:YAG ps pulsed laser in 1987 [35]. In 1993, Tahara and Hamaguchi demonstrated fluorescence background suppression during TR Raman measurements of benzophenone, using an ICCD-based streak camera and a 532 nm Nd:YAG laser in a time-resolution of 10 ps [36]. Moon *et al* used a ps TG Raman spectrometer with pulsed 532-nm excitation in 1994 to reconstruct a 3D image through a turbid medium (water) by amplification, imaging it to a CCD detector [37]. The optical Kerr effect in combination with TG Raman was first described by Deffontaine *et al* in 1985. In their work, the alteration of the refractive index of a Kerr medium (1 mm thick CS₂ cell) in response to a ps laser between crossed polarizers, was applied to perform like a light shutter for TG Raman detection, rather than gating the detector itself [32]. Later in 1999, using a similar ps Kerr gating approach, fluorescence rejection with a more advanced liquid nitrogen-cooled CCD array detector of stepwise 4-(dicyanomethylene)-2-methyl-6-(p-dimethylaminostyryl)-4H-pyran (DCM)-doped acetonitrile was successfully demonstrated by Matousek *et al*, and independently, with a different setup in the ultra-short (fs) range, by Yoshizawa and Kurosawa [38, 39]. With their method, the Matousek group realised fluorescence rejection at variable gating times up to a remarkable minimum of one ps, while Yoshizawa and Kurosawa were actually able to achieve results in the subpicosecond range. However, the Matousek group's setup initially comprised a whole laboratory room, even though they significantly improved the performance of this system in 2001 by using a high-performance ps Kerr gate [38–40]. From the mid-nineties until past the turn of the millennium, various patents have been filed by multiple parties. This includes the US Secretary of Navy's patent relating to TG Raman amplification, involving the use of TG and surface-enhanced Raman spectroscopy (SERS) and TG imaging, as performed by the Massachusetts Institute of Technology (MIT) [41, 42].

Several research groups have studied TG-SERS, including Lecomte *et al*, who focussed on TR surface-enhanced resonance Raman spectroscopy to study the electron-transfer dynamics of heme proteins in 1998 [43]. In addition, a good overview of TR resonance RS is provided by Steven Bell [44].

In 2004, Martyshkin *et al* achieved fluorescence rejection in highly fluorescent solid samples (crystalline CaWO₄ with Nd³⁺ and hexabenzocoronane), with a temporal resolution of approximately 150 ps, using a 2 ps fast pulsed 785 nm Ti:sapphire laser excitation source in combination with an intensified gated/modulated CCD. A similar approach, but with resonance Raman was taken by Efremov *et al* in 2007 to quench the fluorescence from a flavoprotein sample in the UV-region, using cooled ICCD combined with a blue ps excitation source. Figure 4 provides a short summary in the form of a timeline, featuring selected articles described above, together with the

technological drivers for TG Raman, such as subnanosecond pulsed laser systems as key component for TG Raman, as well as different kind of detector variants applicable to the collection of Raman photons in gated mode.

As shown in figure 4, a major boost to the development of TG Raman came with the advent of CMOS SPADs after turn of the millennium. Later, in 2011, CMOS SPAD technology was applied for the first time as a detector technology in Raman applications [20, 45].

2. Fluorescence suppression methods and classification of TG

Several methods have been developed to perform fluorescence rejection. A comprehensive overview is provided by Wei and co-authors [17]. They classify fluorescence interference reduction methods into five different categories: (1) time-domain, (2) frequency-domain, sometimes referred to as 'phase-modulation' under high-frequency modulation, (3) wavelength-domain, (4) computational methods for background removal and post-processing of measured spectra, and (5) other methods. TG is classified as a time-domain method. The first two methods have the major advantage of taking the different temporal behaviour of Raman scattering and fluorescence response into account. The third method utilizes the fact that fluorescence has broader spectral characteristics even at different wavelengths, whereas the Raman emission spectrum is coupled to the wavelength of the excitation [46, 47]. Notable techniques for this method include shifted excitation Raman difference spectroscopy (SERDS), and subtracted shifted Raman spectroscopy, both of which require an additional step, subsequent to spectral acquisition [17]. Only minimal modification is required to convert a conventional CW Raman system into a SERDS setup with CCD based spectrometers, i.e. incorporating two slightly wavelength shifted laser excitation sources, typically set apart at full width at half maximum (FWHM) [48].

Computational methods have a limited capacity to improve the SNR once fluorescence has broadened or distorted Raman peaks. Another drawback is that they can omit sharp Raman peaks, due to random variations in pixel-to-pixel sensitivity that are greater than the actual Raman signal. A significant advantage is that they can be used for baseline-correction, due to the difference between the very narrow Raman peaks as compared to the broad fluorescence [49]. This approach might fail where samples show wider Raman peaks of several tens of wavenumbers. In addition, in some cases, automated computational baseline-correction with single parameters may not be a good fit for all application and sample types, as (i) the risk of data loss and spectral distortion and (ii) manual parameter modification may be required [10, 50]. Other methods for effective fluorescence suppression include sample photobleaching prior to measurement [51], and SERS. A significant enhancement of the Raman scattering can be accomplished using SERS, which increases the Raman scattering when specimens are close (i.e. a distance of nanometres)




Figure 4. Timeline of the development of time-gated Raman spectroscopy, with technological drivers (with ps-pulsed laser shown in green, PMTs and MCPs in yellow, CCDs incl. Kerr-gating in red, and CMOS SPADs in blue), and relevant selected articles.

to metallic surfaces or nanoparticles of matching size and excitation wavelength, creating a plasmonic local electromagnetic enhancement effect to boost the Raman signal, thereby minimising the influence of fluorescence [5]. SERS combines the specificity of Raman with a high sensitivity which enables the analysis of samples at very low analyte concentration, and can also be combined with TG [52]. A more detailed discussion regarding TG-SERS is included in Section 5.

Table 1. Shows a survey of the most notable methods of fluorescence suppression.

Other non-linear techniques, such as coherent anti-Stokes Raman scattering spectroscopy (CARS), can also be classified as fluorescence suppression methods, having the capacity for spatial discrimination of the Raman signal from background interference. CARS has been used for TR measurements in the ps scale, with the aim of rejecting the background from Raman measurements. However, for practical reasons, it is often not applicable to all sample states [22, 53]. Furthermore, the same surface-enhancement methods of boosting Raman signals and suppressing fluorescence can be used for both anti-Stokes Raman and for Stokes (more common) Raman, with the specific advantage in the UV spectral range that vibrations from resonant groups can be selectively picked out [54–56].

3. Gated detector technologies

3.1. PMTs and MCPs

In the 1940s, PMTs were used for the first time as photosensitive low light detectors for Raman experiments [57, 58]. Gating could only be performed by external trigger pulsing, for which in the 1960s PMTs were used to gate stimulated light scattering discrimination, paving the way for future TG Raman detectors [59, 60]. Later MCPs facilitated TG measurements up to the femtosecond range [61, 62]. In this kind of detection arrangement, an image intensifier is placed ahead of photodiode arrays using a microchannel plate. Linearity problems with image intensifiers limited their applicability in combination with the TG measurement setup. A further enhancement in terms of sensitivity can be achieved with intensified photodiode arrays [63]. MCPs are, in principle, electron multipliers in a vacuum-tube assembly, which multiply incident charges to a secondary emission [64]. Since there are many channels permitting spatial resolution, MCPs can be used to resolve a temporal delay [62]. They are also capable of fast switching in the MHz region, making them suitable for TG-related Raman measurements [34]. More commonly, microchannel-plate photomultipliers (MCP-PMTs) are used, as they combine

the advantages of both detector elements. PMTs are special vacuum glass-sealed electron tubes, designed to enhance weak light signals (up to single photons) by generating electrical signals from a photocathode [57, 65].

One drawback of MCP-PMTs is a serious ‘aging’- issue, caused by the ions of residual gases striking and destroying the photo cathode. This results in a rapid deterioration in the quantum efficiency of the detectors, as well as irritating secondary bumps and irregular tails at the FWHM of the instrument response function [66]. This can be addressed by atomic layer deposition and coating with a thin alumina or magnesia layer so as to reduce out-gassing from the MCP substrates [65]. Even though MCP-MPTs seem to be outmoded as Raman detectors, their sensitivity is satisfactory, with suitable temporal resolution, and their development has continued in relation to other applications, as shown in table 2. For example, recent advances have shown that MCP-MPTs make suitable detectors for fluorescence lifetime imaging [67].

3.2. CCDs and ICCDs

In general, CCDs are the most commonly used detector variant for RS, but for the TG setting they need to be highly sensitive (single photon counting ability), allow for a fast external trigger, and possess a temporal resolution in the subnanosecond range. ICCDs fulfil these requirements. Optical Kerr-gating, which acts like a light shutter in front of the entrance slit of the spectrometer, has been used by several groups to trigger the CCD. This setup requires space, thus limiting the portability of the system. Guidelines for the selection of Raman multichannel and gated detection have been formulated by Talmi [71]. In 1993, Tahara and Hamaguchi were the first to achieve high sensitivity and good timing resolution by constructing an intensified CCD-based streak camera. The streak camera in a TG Raman setup guides the backscattered light from the sample onto a photocathode; this accelerates electrons when hit by photons, by means of high speed voltage ramps (for positive and negative DC bias) on an anode in a cathode tube (known as a streak tube). The movement of the electron beam (streak) is swapped from the negative anode side to the positive, before being passed through a phosphor screen via an intensifier onto the CCD detector. This mode of sweeping enables a detection result for a three-dimensional plot (intensity vs. frequency and time) [72].

An extended option for performing streak-based TG measurements with increased detection sensitivity is the use of electron-multiplying charge-coupled devices (EMCCD), which is an additional feature of some CCD detectors [73–75].

ICCDs often require Peltier cooling (from -20°C to below -100°C), which is not generally water-cooling or liquid nitrogen cooling. These components affect the electronics, overall size, cost, and complexity of ICCDs, which are still relatively sizeable at present.

3.3. SPADs

The CMOS industrial technology standard enabled the design of high-speed electronics in combination with SPADs at low

cost and in smaller sizes, without the need for detector cooling. This made it possible to arrange simplified photo-detection and fast (subnanosecond) readout circuitry on a tiny solid state chip [19].

The very first SPAD array detector was developed by Rochas *et al* in 2003 [76]. SPAD detectors are very sensitive, produce no significant readout noise at high speeds, allow for precise timing, and do not require additional preamplifiers for this level of sensitivity, in contrast to CCDs [68]. Even a single photon entering the depletion region on the detector can trigger an avalanche breakdown that can directly produce high-speed logic-level pulses (bias) for the read-out measurement electronics [9]. Generally speaking, a SPAD is an avalanche photodiode, in which the p-n junction is reverse biased above its breakdown voltage, operating in the so-called ‘Geiger Mode’. When a photon hits the active area, it creates an electron-hole pair, triggering an avalanche [9, 20, 68].

3.4. Advantages and disadvantages of SPADs for TG Raman vs other detector techniques

The advantages of SPAD detectors over MCPs, PMTs and CCDs for TG Raman can be summarized as follows: (1) greater suitability for TG, due to their temporal behaviour, (2) advanced sensitivity, since SPADs operate close to the ‘break-down’ voltage in Geiger mode, thereby enabling single photon detection and the exposure of fast transitions at very low intensities, (3) less complexity, resulting in lower manufacturing costs, (4) the capacity for improved system integration, resulting in reduced space requirements, (5) less power consumption/improved power dissipation and (6) no cooling requirements, due to the detectors’ short duty-cycle and low dark current [10, 17, 77, 78]. One current disadvantage of SPAD detectors in relation to TG Raman applications is the challenge of matching a comparable number of pixels in the detector array, as compared to CCDs. This can have an effect on the spectral resolution, although there are ways to improve this, e.g. the implementation of microlens arrays and sub-pixel acquisition [68, 79, 80]. Current commercial TG Raman spectrometers provide a spectral resolution of about $5\text{ (cm}^{-1}\text{)}$ wavenumbers, while some CCD-based systems can achieve below $1\text{ (cm}^{-1}\text{)}$ [25, 81]. However, the majority of applications do not require sub-wavenumber resolutions.

3.5. Overview of SPAD-detector development for TG Raman

The first applications using SPAD technology for TG RS were demonstrated by Blackberg *et al*, and by Nissinen *et al*, in 2011 [20, 45]. The Nissinen group used the rising edge of a 300 ps pulsed Nd:YAG micro-chip laser, at a 532 nm excitation wavelength, to trigger a delay generator and timing circuits to enable the SPAD, detecting the collected Raman photons on one SPAD element. Later in 2013, Kostamovaara *et al*, using a similar setup, demonstrated that gating times of around 100 ps are sufficient for most sample-induced fluorescence suppression schemes [9]. The early setups used a single pixel SPAD element and a translational stage that moved the SPAD detector over the spectrograph’s output slit in order to resolve

Table 2. Comparison of TG Raman-compatible detector technologies, compiled from references [24, 68–70].

Detector type	Multichannel plates (MCPs)	Photo-multiplier tubes (PMTs)	MCP-MPTs	Intensified gated charge-coupled devices (ICCDs)	Complementary metal oxide semiconductor single-photon avalanche diodes (CMOS SPADs)
Instrument complexity	High	High	Medium	Medium to high	Low
Size	Large	Large	Large to medium	Medium	Medium/compact/portable
Sensitivity	Good	Good	Good	Good	Very good
Time-gate range depending on excitation source	Femtoseconds	Nanoseconds	Picoseconds, up to Femtoseconds, depending on setup	Picoseconds	Picoseconds
Necessity for cooling	–	Beneficial	Beneficial	Required	Not required
Cost	–	–	–	High	Medium
Remarks	Laboratory setup for TG Raman	Laboratory setup for TG Raman; can be used with optical Kerr gates	Laboratory scale; can be used with optical Kerr gates	Typically, larger setup required. Can be used with optical Kerr gates. Streak camera and EMCCD options available. Commercial availability	Original application as single-photon counting devices. Commercial availability, e.g. as portable process Raman spectrometer instruments

a full Raman spectrum. There are multiple variants of two-dimensional SPAD array detectors suitable for TG Raman applications, as initially demonstrated in 2013 by Nissinen *et al* [82]. A detailed overview of CMOS SPADs for biophotonics applications is provided by Bruschini *et al* 2019 [68]. Details of recent technical advances in CMOS SPAD detectors for TG RS were provided by Nissinen *et al's* 2017 paper [19]. They demonstrate a compact (9 mm × 3 mm) 16 × 256 CMOS SPAD line sensor with low power consumption, and a 256-channel time-to-digital converter (TDC) on the chip. The 3-bit TDC is placed near the SPAD array, facilitating an adjustable timing resolution from 52–104 ps. This is manufactured via a 0.35 μm high-voltage CMOS process, similar to Maruyama *et al's* 2014 8 × 1024 SPAD-array, with less active pixels, but a greatly improved FWHM, an increased fill factor, and a smaller time-gate width [69, 83]. Detailed comparisons of CMOS SPADs for single-photon detection, and for Raman applications in particular, can be found elsewhere [69, 80, 84].

4. Application areas and future directions

As mentioned above, pulsed laser techniques enable TG and TR measurements. However, application areas for these techniques might differ, depending on whether the aim relates purely to background (e.g. fluorescence) suppression or to transient measurements, i.e. studying the temporal emission distribution. Based on our findings in the literature, TG has successfully been used in several application areas. These findings enable us to estimate some of the most suitable application areas for TG Raman, as shown in figure 5. RS is an important online-monitoring tool with the capability for quantitative and qualitative analysis in process industrial sensing

applications; currently its main fields of application in relation to TG Raman are the steel industry or industrial catalysis [85, 86]. These fields benefit in particular from the capacity of RS to perform a real-time and *in situ* non-invasive and non-destructive analysis at different locations in a process, with zero or minimal sample preparation (e.g. flow-cell or reaction chamber), together with its capacity to identify relevant structural information and polymorphism with high levels of spectral and spatial composition for compounds at different process phases [86–92]. All this is possible with fluorescent-free and undisturbed Raman detection [93]. Notable in terms of process analytical applications is the fact that gated Raman measurements can also be conducted in transmission mode, i.e. through turbid media (using flow-cells or pipes), with improved SNR by means of spatially offset Raman spectroscopy [94, 95]. TG Raman industrial applications include the following published examples: Heilala *et al* investigated traces of acids in pickle liquor in metalworking industry. Gyakwaa *et al* successfully characterized Calcium-Aluminate in steel production, as steel and most other alloys produce minimal Raman intensities and thus unwanted inclusions and other components can be detected [85, 86]. Gal-Or *et al* demonstrated the feasibility of using TG Raman in 3D printing manufacturing of glass microfluidics [96], and Dou *et al* confirmed, using TG Raman and NMR spectroscopy, that highly fluorescent willow bark is a rich source and a potential feedstock for biorefinery processes in fibre production [97]. Catalytic process applications using where TG/TR Raman performs an online-monitoring function in relation to transition metal oxides [92] and the optimization of SiO₂—TiO₂ concentration ratios at specific temperatures has also been studied. Enzyme biocatalytic reaction processes have also shown promising results using this type of monitoring technology




Figure 5. Applications of time-gated Raman found in literature (related to the evaluated articles) ranging from process industrial sensing to hot temperature and environmental sensing.

[98]. Furthermore, qualitative and quantitative Raman process measurements are mostly unaffected by vibrations and high pressure caused by high flow-rates or generated through pumps in the pathway, so long as the detection takes place on a fixed sample substrate, e.g. membranes in water filtration processes [99]. The effect of changes in pH and temperature (e.g. hot temperature and catalysis processes) on Raman and IR spectra manifest as small band drifts, but they are not negatively influenced by e.g. very large drifts or intensity changes in bands, which is a great advantage when following process-related changes, and is even more suitable for pulsed laser and TG or TR measurements in general [100–103]. Sharma *et al* have demonstrated excellent results for fluorescence suppression using pulsed laser TR Raman applied in NASA's Venus Chamber, at temperatures of up to 1373 K and at high pressures [102]. Promising results aimed at planetary sciences, but also at more general mineralogical sciences, have been realized using TG/TR Raman to investigate specimens of minerals with very strong fluorescence interference [46, 104]. These results include the investigation of rock samples containing rare earth elements [93, 105, 106].

A recent focus of TG Raman in the biomedical arena has been the chemical imaging of human teeth, as described by Kekkonen *et al* [107], and Usai *et al* [108], who compared the results of samples from pure distilled-water fluorescein, paracetamol, and sesame oil, known to have high fluorescence backgrounds, with TG (uncooled 1024×8 SPAD line detector). Compared with the results obtained using conventional CW Raman (cooled CCD), at 532 nm excitation, they were able to demonstrate remarkable fluorescence suppression results using TG. Knorr *et al* developed a CCD-based TG Raman system to study biofuels [109]. They noticed that

high laser peak power at low repetition rates caused sample surface damage, but that sufficient spectra can be measured below a certain laser power threshold. Morris and co-authors used ps TG Raman for the transcutaneous evaluation of bone compositions, using the Kerr gate to study bone compositions from below the surface (phosphates around $945\text{--}957\text{ cm}^{-1}$, carbonates 1073 cm^{-1} , amide II region at 1250 cm^{-1} , CH_2 at 1445 cm^{-1} , and amide I envelope at 1685 cm^{-1}). Transcutaneous Raman bands are observable from 0–10 ps delay, without fluorescence interference and sufficient SNR, but increased fluorescence and decreased SNR. Further studies on living animals or human subjects are proposed, where scanning and real-time detection of a larger tissue surface area may be possible [110]. Rojalín *et al* demonstrated fluorescence suppression with a CMOS SPAD TG Raman prototype, based on pharmaceutical samples that produced a high degree of fluorescence interference, such as caffeine, ranitidine, and indomethacin powders [21]. A residual fluorescence background (REL) subtraction was implemented, together with a decay estimation function, which is currently used in some commercial TG Raman spectrometers [25]. Conventional 785 nm CW Raman and 532 nm TG Raman measurements for these samples were compared, revealing that TG could significantly dampen the fluorescence background. Lipiäinen *et al* applied multivariate chemometric data analysis to solid fluorescent pharmaceutical drug mixtures, using a commercially ready CMOS SPAD prototype system. A precise estimation of the compound quantification of the mixtures was achieved, with partial least-square (PLS) and kernel-based regularized least-squares regression (RLS), employed where a large amount of spectral data is considered for Raman signal intensity due to overlapping peaks. Chemometric analysis using PLS and RLS was conducted on data acquired using a CMOS SPAD prototype TG Raman, and the individual temporal intensity decays of each mixture's fluorescence was also included in the study.

In the field of biotechnology, Kögler *et al* showed in 2018 that TG Raman, accomplished using a commercially ready CMOS SPAD prototype system with and w/o nanoparticle enhancement, could identify significantly more Raman and SERS bands than the common CW process Raman spectrometers and confocal Raman microscopes in relation to *Escherichia coli* (*E. coli*) cultivation samples over a 24 h process time [111]. In addition to the qualification of relevant compounds such as glucose and acetate, data detailing changes in nucleotides (DNA) taken from TG Raman and CW Raman were chemometrically quantified and compared with high-performance liquid chromatography. Kögler *et al* concluded that even with a spectral resolution of 10 cm^{-1} , the SPAD-based TG Raman is more suitable for qualification and quantification than CW Raman, due to its fluorescence suppression capability, and its combined plasmonic enhancement with SERS. Further experiments, without the use of plasmonic enhancement, on yellow fluorescent protein (YFP), cultivated over an extended process time, and using the same mineral salt medium as for the *E. coli* studies, are illustrated in figure 6. Data obtained from TG Raman were taken at the beginning of the cultivation period (after approximately 1 h (blue curve))




Figure 6. Bioprocess application—comparison of (a) TGRaman (at $\lambda_{\text{exc}} = 532$ nm) and (b) CWRaman (at $\lambda_{\text{exc}} = 785$ nm) for yellow fluorescent protein (YFP) in a turbid growth medium with strong fluorescence interference during cultivation at 1 h cultivation time (blue curve) and towards the end at 9 h cultivation time (red curve), modified and reprinted with permission from reference [10].




Figure 7. TG-SERS (picosecond pulsed excitation) versus CW-SERS (continuous excitation) at $\lambda_{\text{exc}} = 532$ nm—(a) *E. coli* cells in complex media expressing hCNTF (1 h after induction) and (b) eGFP in buffer—TG-SERS (green curve) and conventional CW-SERS (red curve)—reprinted with permission from reference [23].

and at its end (after about 9 h (red curve)), as shown in figure 6(a). These were then compared with data obtained via CW Raman (figure 6(b)). The results clearly indicate that TG Raman identifies significantly more known Raman bands for this process than CW Raman, and that fluorescence causes distortion in the Raman signature (only one visible Raman peak at around 980 cm^{-1} with $\lambda_{\text{exc}} = 785$ nm CW Raman) with a broad interference from 1000 cm^{-1} onwards [10].

For process analytical applications, and for most application areas mentioned in figure 5, a TG Raman spectrometer needs to be a high-performance device; it should be portable and compatible with different kind of probe options for specific applications. The device size and portability are also important factors in many applications, although they might be less crucial when monitoring mature processes. Modifications to reduce costs and size to meet specific application needs is possible with TG Raman. As an example, by 2014, Li and Deen had already developed a portable TG Raman system, based on CMOS SPAD technology, intended to be used in field and environmental applications, particularly for water quality monitoring in remote areas [73, 84]. They stated that the

largest component in the system is the spectrograph, which can be downsized (approx. 2.5×4 cm in their system) by employing a concave grating design [84]. However, future directions for RS applications aim to achieve miniaturization down to the size of handheld devices, or even smartphone-based systems [112, 113]. These portable systems may be very suitable for environmental sensing, mineral & rock identification, counterfeit detection, and bio-sample detection in the field. We believe that an even greater degree of miniaturization in TG Raman systems is possible. One example of a portable TG Raman system is the SuperCam Instrument, carried on NASA's Mars 2020 rover, which includes a TR pulsed laser Raman spectrometer [114].

5. Extensions and further developments of TG Raman

Some techniques can be seen as extensions of TG Raman, because the TG mode enhances the capabilities of these techniques far beyond what they are capable of without TG.




Figure 8. Standoff TGRaman, based on a compact time-correlated complementary metal-oxide-semiconductor (CMOS) single-photon avalanche diode (SPAD) array sensor and ps pulsed laser, reprinted with permission from reference [128].

SERS and standoff Raman detection may be viewed as such extensions to TG, with great potential for further development in the future.

5.1. TG-SERS

SERS requires suitable metallic nanostructures or planar electrode surfaces (typically silver or gold) in order to achieve plasmonic signal enhancement in Raman technology. Noble metallic nanoparticles of specific size and shape or specially patterned nanostructures are suitable; sometimes combination of both seems applicable [115, 116]. SERS has the advantage of strongly enhancing the Raman signal and, like Raman, is insensitive to the molecular vibrations of water [117, 118]. Fluorescence can be partially quenched via plasmonic enhancement, but this is dependent on the quantum yield of the molecules under inspection [5, 10]. Even low-end specification Raman spectrometers can be employed with very reasonable results; thus, SERS does not require a high-end process or confocal Raman spectrometer systems [119]. Surface enhancement is based on the resonance excitations of surface plasmons in the metal, and enhancement factors of about 10^4 – 10^6 , in extreme cases even about 10^{11} , can be achieved [5]. SERS is currently by far the most predominant subject in the field of RS research, but due to the lack of long-term repeatability in terms of spectral results, and variable performance, it has yet to make a meaningful breakthrough in the industry [10]. More in depth information about SERS can be found elsewhere [120–122].

While TG helps to overcome the issue of fluorescence and other background interference, it cannot improve the generally weak Raman scattering phenomenon in relation to the strong

fluorescence signal, as discussed in the introduction. Improving the limit of detection is highly desirable, particularly for low analyte concentrations. Another related aspect is the aforementioned ability of SERS to quench fluorescence to some extent. Conceivably, therefore, a TG-SERS combination may help to overcome even a very strong influence of fluorescence, as Kögler *et al* have demonstrated [111, 116].

Most recently, Kögler *et al* assessed the capacity of TG-SERS to reduce high fluorescence signals from complex cultivation media of *E. coli* cells in a bioprocess. They also compared the results obtained from TG-SERS with the non-enhanced evaluation of TG and CW Raman spectra [23].

Figure 7 illustrates the difference between TG-SERS (green spectra) and common CW-SERS (red spectra) results, both at $\lambda_{exc} = 532$ nm, in comparison with the same sample from a bioprocess. Figure 7(a) shows the results for an *E. coli* sample at the beginning of its cultivation period, and figure 7(b) gives results for an enhanced green fluorescent protein (eGFP) with an even stronger background fluorescence.

The spectral results shown in figure 7 are from highly fluorescent samples, but at relatively low concentration levels; thus, a simultaneous Raman enhancement and fluorescence suppression with TG-SERS is beneficial for these applications.

5.2. TG standoff Raman detection

Remote Raman analysis of dangerous (e.g. explosives or bio-hazards), fragile, mineral or inaccessible samples can be performed at a safe standoff distance of up to several hundred metres. A good overview of this method and its applications can be found elsewhere [123, 124]. For example, standoff Raman detection has been used to detect pollutants in air and

water at great distances, as well planetary exploration for the detection of organics mixed in mineral matrices [125–127]. The reason why the TG principle plays an important role in this method stems from the need to reduce the disturbing amount of fluorescence and the background radiation of ambient light from the often unknown heterogeneous sample matrices under inspection. Often, a pulsed laser system around the fluorescence maximum at $\lambda_{\text{exc}} = 532$ nm, or in the UV-region, with very sensitive and gated detectors (ICCDs or SPADs), is employed for this type of application [128, 129]. Recently, Kekkonen *et al* (2018) proposed CMOS SPAD detectors for remote Raman detection to further reduce size, price and complexity, with an additional adjustable time-correlation option, for a TG standoff Raman setup. They demonstrated remote distance Raman results for TiO_2 using an uncooled low power consumption 16×256 CMOS SPAD line sensor, with a 256-channel TDC, and a high scanning resolution of 3.75 cm, at different target distances. Figure 8 shows the TG standoff Raman setup with the CMOS SPAD array detector. It should be noted that the measurement distance to the target can be increased with a telescope aligned with the lens, as shown in figure 8.

Furthermore, an important role for standoff Raman is in terms of its application in the *in situ* detection of explosives [130]. Gaft and Nagli evaluated already in 2007 gated-Raman for standoff detection of explosives (factory and homemade). They noticed that the binder in explosives creates a strong luminescence signal, and that near-infrared Raman seems to be inapplicable for the detection of explosives, given the presence of strong background fluorescence. Only a stronger scattering probability in UV can overcome the fluorescence issue, since many explosives have maximal absorption bands in parts of the UV spectrum. TG UV Raman excitation can minimize fluorescence, and increases the Raman signal by a factor of about 100–200 [49].

6. Conclusion and outlook

Since the discovery of Raman fluorescence background suppression in the 1970s by Yaney, using a TG PMT detector setup in combination with a pulsed excitation laser source, many new technical approaches, application areas and extensions of this method have emerged. Despite the option to enhance the Raman signal with plasmonic nanostructures (SERS), the disturbing sample-induced fluorescence background still seems to be a major issue across all Raman-related disciplines. Even SERS could benefit from further fluorescence suppression, by combining it with TG Raman. Different detector techniques, ranging from MCP/PMTs, ICCDs with and without Kerr gates, and CCD-based streak cameras, to efficient and compact CMOS SPAD arrays have been used to perform TG Raman. High laser pulse intensities to obtain a sufficient SNR are no longer necessary, due to the increased sensitivity, larger number of active pixels, and improved fill factor available in most recent SPAD-array detectors.

TG Raman technology has in the past been criticised due to its requirements for very complex, expensive and large instrumentation. This problem has been overcome thanks to advances in the CMOS process, which have enabled the fabrication of RS-dedicated SPAD technology. Furthermore, affordable small and fast enough picosecond pulsed micro-chip laser at 532 nm excitation have only recently become available. As the excitation and detector side of the technology matures, it is also possible to move to other regions in the emission spectrum to further reduce auto-fluorescence. Commercial high-performance TG Raman spectrometers for process analytical applications are now available. These are portable and compact in size. Less highly performing but even smaller TG Raman spectrometers would open up more application areas, and have been already demonstrated for various field and environmental sensing applications. A possible future direction is towards mobile and handheld Raman devices, using a TG Raman spectrometer to connect to a smartphone, able to verify measured results via a spectral library in the cloud.

Acknowledgments

We thank the sensing and integration department of VTT, the technical research centre of Finland, for their support in the production of this review article.

Conflict of interest

Co-author Bryan Heilala is affiliated with the company which commercialized the CMOS SPAD detector technology described in this work. The first author has no conflict of interest.

ORCID iDs

Martin Kögler  <https://orcid.org/0000-0003-2154-8879>

Bryan Heilala  <https://orcid.org/0000-0002-2738-1259>

References

- [1] Breed G 2010 Fundamentals of pulsed and time-gated measurements *High Freq. Electron.* **9** 1–4
- [2] Panasenko D, Sun P-C, Alic N and Fainman Y 2002 Single-shot generation of a sonogram by time gating of a spectrally decomposed ultrashort laser pulse *Appl. Opt.* **41** 5185
- [3] Van Duyne R P, Jeanmaire D L and Shriver D F 1974 Mode-locked laser Raman spectroscopy. A new technique for the rejection of interfering background luminescence signals *Anal. Chem.* **46** 213–22
- [4] Lewis I R and Edwards H 2001 *Handbook of Raman Spectroscopy: From the Research Laboratory to the Process Line* (Boca Raton, FL: CRC Press Taylor & Francis)
- [5] Procházka M 2016 *Surface Enhanced Raman Spectroscopy* (Berlin: Springer)
- [6] van de Hulst H C 1981 *Light Scattering by Small Particles* (New York: Dover Publications Inc.)
- [7] Smith E and Dent G 2005 *Modern Raman Spectroscopy* (New York: Wiley)

- [8] Esmonde-White K A, Cuellar M, Uerpmann C, Lenain B and Lewis I R 2016 Raman spectroscopy as a process analytical technology for pharmaceutical manufacturing and bioprocessing *Anal. Bioanal. Chem.* **409** 637–49
- [9] Kostamovaara J, Tenhunen J, Kögler M, Nissinen I, Nissinen J and Keränen P 2013 Fluorescence suppression in Raman spectroscopy using a time-gated CMOS SPAD *Opt. Express* **21** 31632
- [10] Kögler M 2018 *Advanced Raman Spectroscopy for Bioprocess Monitoring* (Technical University of Berlin) Berlin
- [11] Kudryavtsev D A, Fedotenko T, Koemets E G, Khandarkhaeva S E, Kutcherov V G and Dubrovinsky L S 2020 Raman spectroscopy study on chemical transformations of propane at high temperatures and high pressures *Sci. Rep.* **10** 1–10
- [12] Lefebvre J, Maruyama S and Finnie P 2007 Photoluminescence: science and applications Jacques *Carbon Nanotubes*, ed A Jorio, G Dresselhaus and M S Dresselhaus (Berlin: Springer) pp 287–319
- [13] Mihoková E and Nikl M 2015 Luminescent materials: probing the excited state of emission centers by spectroscopic methods *Meas. Sci. Technol.* **26**
- [14] Birch D J S and Imhof R E 2002 Time-Domain Fluorescence Spectroscopy Using Time-Correlated Single-Photon Counting *Topics in Fluorescence Spectroscopy* vol 1 (Boston: Springer) pp 1–95
- [15] Kneipp J, Kneipp H and Kneipp K 2008 SERS—a single-molecule and nanoscale tool for bioanalytics *Chem. Soc. Rev.* **37** 1052–60
- [16] Ekins S 2008 *Pharmaceutical Applications of Raman Spectroscopy* (New Jersey: Wiley)
- [17] Wei D, Chen S and Liu Q 2015 Review of fluorescence suppression techniques in Raman spectroscopy *Appl. Spectrosc. Rev.* **50** 387–406
- [18] Yaney P P 1976 The pulsed laser and gated detection in Raman spectroscopy – a survey of the spectra of common substances including studies of adsorbed benzene *J. Raman Spectrosc.* **5** 219–41
- [19] Nissinen I, Nissinen J, Keränen P and Kostamovaara J 2017 On the effects of the time gate position and width on the signal-to-noise ratio for detection of Raman spectrum in a time-gated CMOS single-photon avalanche diode based sensor *Sensors Actuators B* **241** 1145–52
- [20] Nissinen I, Nissinen J, Lämsman A K, Hallman L, Kilpelä A, Kostamovaara J, Kögler M, Aikio M and Tenhunen J 2011 A sub-ns time-gated CMOS single photon avalanche diode detector for Raman spectroscopy: *Eur. Solid-State Device Res. Conf.* pp 375–8
- [21] Rojalín T, Kurki L, Laaksonen T, Viitala T, Kostamovaara J, Gordon K C, Galvis L, Wachsmann-Hogiu S, Strachan C J and Yliperttula M 2016 Fluorescence-suppressed time-resolved Raman spectroscopy of pharmaceuticals using complementary metal-oxide semiconductor (CMOS) single-photon avalanche diode (SPAD) detector *Anal. Bioanal. Chem.* **408** 761–74
- [22] Howard J, Everall N J, Jackson R W and Hutchinson K 1986 Fluorescence rejection in Raman spectroscopy using a synchronously pumped, cavity-dumped dye laser and gated photon counting *J. Phys. E.* **19** 934–43
- [23] Kögler M, Itkonen J, Viitala T and Casteleijn M G 2020 Assessment of recombinant protein production in *E. coli* with Time-Gated Surface Enhanced Raman Spectroscopy (TG-SERS) *Sci. Rep.* **10** 1–11
- [24] Kekkonen J, Talala T, Nissinen J and Nissinen I 2020 On the spectral quality of time-resolved CMOS SPAD-based Raman spectroscopy with high fluorescence backgrounds *IEEE Sens. J.* **PP** 10
- [25] Timegate Instruments Oy F 2020 PicoRaman
- [26] Wirth M J and Chou S H 1988 Comparison of time and frequency domain methods for rejecting fluorescence from Raman spectra *Anal. Chem.* **60** 1882–6
- [27] Blacksborg J, Alerstam E, Cochrane C J, Maruyama Y and Farmer J D 2020 Miniature high-speed, low-pulse-energy picosecond Raman spectrometer for identification of minerals and organics in planetary science *Appl. Opt.* **59** 433–44
- [28] Blacksborg J, Rossman G R and Gleckler A 2010 Time-resolved Raman spectroscopy for in situ planetary mineralogy *Appl. Opt.* **49** 4951–62
- [29] Blacksborg J, Alerstam E, Maruyama Y, Cochrane C J and Rossman G R 2016 Miniaturized time-resolved Raman spectrometer for planetary science based on a fast single photon avalanche diode detector array *Appl. Opt.* **55** 739–48
- [30] Yaney P P 1972 Reduction of fluorescence background in Raman spectra by the pulsed Raman technique. *J. Opt. Soc. Am.* **62** 1297–303
- [31] Harris J M, Chrisman R W, Lytle F E and Tobias R S 1976 Sub-nanosecond time-resolved rejection of fluorescence from Raman spectra *Anal. Chem.* **48** 1922–6
- [32] Deffontaine A, Delhaye M and Bridoux M 1985 Pulsed multichannel Raman technique *Rev. Phys. Appl.* **19** 415–21
- [33] Watanabe J, Kinoshita S and Kushida T 1985 Fluorescence rejection in Raman spectroscopy by a gated single-photon counting method *Rev. Sci. Instrum.* **56** 1195–8
- [34] Everall N, Jackson R W, Howard J and Hutchinson K 1986 Fluorescence rejection in Raman spectroscopy using a gated intensified diode array detector *J. Raman Spectrosc.* **17** 415–23
- [35] Petrich J W, Martin J L, Houde D, Poyart C and Orszag A 1987 Time-resolved Raman spectroscopy with subpicosecond resolution: vibrational cooling and delocalization of strain energy in photodissociated (carbonmonoxy)hemoglobin. *Biochemistry* **26** 7914–23
- [36] Tahara T and Hamaguchi H O 1993 Picosecond Raman spectroscopy using a streak camera *Appl. Spectrosc.* **47** 391–8
- [37] Moon J A, Mahon R, Duncan M D and Reintjes J 1994 Three-dimensional reflective image reconstruction through a scattering medium based on time-gated Raman amplification *Opt. Lett.* **19** 1234
- [38] Yoshizawa M and Kurosawa M 2000 Femtosecond time-resolved Raman spectroscopy using stimulated Raman scattering *Phys. Rev. A - At. Mol. Opt. Phys.* **61** 6
- [39] Matousek P, Towrie M, Stanley A and Parker A W 1999 Efficient rejection of fluorescence from Raman spectra using picosecond Kerr gating *Appl. Spectrosc.* **53** 1485–9
- [40] Matousek P, Towrie M, Parker A W, Facility C L, Rutherford C, Ma C, Kwok W M and Phillips D 2001 Fluorescence suppression in Raman spectroscopy using a high performance picosecond Kerr gate *J. Raman Spectrosc.* **32** 983–8
- [41] Perelman L T, Wu J, Wang Y, Dasari R R, Itzkan I and Feld S M 1999 Optical imaging using time gated scattered light *US Patent Specification* 5,837,964 19
- [42] Reintjes J F, Duncan M D, Mahon R, Tankersley L L, Waynant R W and Bashkansky M 1994 Time-gated imaging through defense-scattering materials using stimulated Raman amplification *US Patent Specification*
- [43] Lecomte S, Wackerbarth H, Soulimane T, Buse G and Hildebrandt P 1998 Time-resolved surface-enhanced resonance Raman spectroscopy for studying electron-transfer dynamics of heme proteins *J. Am. Chem. Soc.* **120** 7381–2

- [44] Bell S E J 1996 Time-resolved resonance Raman spectroscopy - tutorial review *Analyst* **121** 107–20
- [45] Blacksborg J, Maruyama Y, Charbon E and Rossman G R 2011 Fast single-photon avalanche diode arrays for laser Raman spectroscopy *Opt. Lett.* **36** 3672
- [46] Böttger U *et al* 2017 Shifted excitation Raman difference spectroscopy applied to extraterrestrial particles returned from the asteroid Itokawa *Planet. Space Sci.* **144** 106–11
- [47] Maiwald M, Müller A, Sumpf B and Tränkle G 2016 A portable shifted excitation Raman difference spectroscopy system: device and field demonstration *J. Raman Spectrosc.* **47** 1180–4
- [48] Sowoidnich K and Kronfeldt H-D 2012 Fluorescence rejection by shifted excitation Raman difference spectroscopy at multiple wavelengths for the investigation of biological samples *ISRN Spectrosc.* **2012** 1–11
- [49] Gaft M and Nagli L 2008 UV gated Raman spectroscopy for standoff detection of explosives *Opt. Mater.* **30** 1739–46
- [50] Guo S, Bocklitz T and Popp J 2016 Optimization of Raman-spectrum baseline correction in biological application *Analyst* **141** 2396–404
- [51] Zięba-Palus J and Michalska A 2014 Photobleaching as a useful technique in reducing of fluorescence in Raman spectra of blue automobile paint samples *Vib. Spectrosc.* **74** 6–12
- [52] Popp J and Mayerhöfer T 2009 Editorial: surface-enhanced Raman spectroscopy *Anal. Bioanal. Chem.* **394** 1717–8
- [53] Ariunbold G and Altangerel N 2017 Quantitative interpretation of time-resolved coherent anti-Stokes Raman spectroscopy with all Gaussian pulses *J. Raman Spectrosc.* **48** 104–7
- [54] Steuwe C, Kaminski C F, Baumberg J J and Mahajan S 2011 Surface enhanced coherent anti-stokes Raman scattering on nanostructured gold surfaces *Nano Lett.* **11** 5339–43
- [55] Hua X, Voronine D V, Ballmann C W, Sinyukov A M, Sokolov A V and Scully M O 2014 Nature of surface-enhanced coherent Raman scattering *Phys. Rev. A - At. Mol. Opt. Phys.* **89** 1–7
- [56] Namboodiri V, Namboodiri M, Diaz G I C, Oppermann M, Flachenecker G and Materny A 2011 Vibrational spectroscopy surface-enhanced femtosecond CARS spectroscopy (SE-CARS) on pyridine *Vib. Spectrosc.* **56** 9–12
- [57] Engstrom R W 1947 Multiplier photo-tube characteristics: application to low light levels* *J. Opt. Soc. Am.* **37** 420
- [58] Rank D H, Pfister R J and Coleman P D 1942 Photoelectric detection and intensity measurement in Raman spectra *J. Opt. Soc. Am.* **32** 390
- [59] De Martini F and Wacks K P 1967 Photomultiplier gate for stimulated-spontaneous light scattering discrimination *Rev. Sci. Instrum.* **38** 866–8
- [60] Burgess S and Shepherd I W 1977 Fluorescence suppression in time-resolved Raman spectra *J. Phys. E.* **10** 617–20
- [61] Deffontaine A, Bridoux M and Delhay M 1984 The third generation of multichannel Raman spectrometers *Rev. Phys. Appl.* **19** 415–21
- [62] Champion A and Woodruff W H 1987 Multichannel Raman spectroscopy *Anal. Chem.* **59** 1299–308
- [63] Antcliff R R, Hillard M E and Jarrett O 1984 Intensified silicon photodiode array detector linearity: application to coherent anti-Stokes Raman spectroscopy *Appl. Opt.* **23** 2369
- [64] Courtney S H and Wilson W L 1991 Multichannel time-correlated single photon counting: spectroscopy and time-gated imaging using a resistive anode photomultiplier tube *Rev. Sci. Instrum.* **62** 2100–4
- [65] Lehmann A *et al* 2020 Latest improvements of microchannel-plate PMTS *Nucl. Instruments Methods Phys. Res. Sect. A Accel. Spectrometers, Detect. Assoc. Equip.* **958** 162357
- [66] Spinelli A, Ghioni M A, Cova S D and Davis L M 1998 Avalanche detector with ultraclean response for time-resolved photon counting *IEEE J. Quantum Electron.* **34** 817–21
- [67] Prokazov Y, Turbin E, Weber A, Hartig R and Zuschratter W 2014 Position sensitive detector for fluorescence lifetime imaging *J. Instrum.* **9** C12015
- [68] Bruschini C, Homulle H, Antolovic I M, Burri S and Charbon E 2019 Single-photon avalanche diode imagers in biophotonics: review and outlook *Light Sci. Appl.* **8** 1–28
- [69] Nissinen I, Nissinen J, Keranen P, Stoppa D and Kostamovaara J 2018 A16 × 256 SPAD line detector with a 50-ps, 3-bit, 256-channel time-to-digital converter for Raman spectroscopy *IEEE Sens. J.* **18** 3789–98
- [70] Pancheri L, Massari N and Stoppa D 2013 SPAD image sensor with analog counting pixel for time-resolved fluorescence detection *IEEE Trans. Electron Devices* **60** 3442–9
- [71] Talmi Y 1988 Guidelines for the selection of raman multichannel detectors *Proc. SPIE* (<https://doi.org/10.1117/12.941947>)
- [72] Baker R J and Johnson B P 1994 Sweep circuit design for a picosecond streak camera *Meas. Sci. Technol.* **5** 408–11
- [73] Li Z 2015 *Miniaturization of Time-Gated Raman Spectrometer with a Concave Grating and a CMOS Single Photon Avalanche Diode* Miniaturization of time-gated Raman spectrometer with a concave grating and a CMOS single photon avalanche diode *Ph.D Thesis*, McMaster University
- [74] Zong C, Chen C J, Zhang M, Wu D Y and Ren B 2015 Transient electrochemical surface-enhanced Raman spectroscopy: a millisecond time-resolved study of an electrochemical redox process *J. Am. Chem. Soc.* **137** 11768–74
- [75] Golcuk K, Mandair G S, Callender A F, Sahar N, Kohn D H and Morris M D 2006 Is photobleaching necessary for Raman imaging of bone tissue using a green laser? *Biochim. Biophys. Acta* **1758** 868–73
- [76] Rochas A, Gani M, Furrer B, Besse P A, Popovic R S, Ribordy G and Gisin N 2003 Single photon detector fabricated in a complementary metal-oxide-semiconductor high-voltage technology *Rev. Sci. Instrum.* **74** 3263–70
- [77] Zappa F, Tisa S, Tosi A and Cova S 2007 Principles and features of single-photon avalanche diode arrays *Sensors Actuators A* **140** 103–12
- [78] Smith J A, Dhulla V H, Mukherjee S S, Lauenstein J-M, Hare R J, Zorn C J and Hostetler C A 2020 Evaluation of an operational concept for improving radiation tolerance of single photon avalanche diode (SPAD) arrays *IEEE Trans. Nucl. Sci.* **9499** 1–1
- [79] Düppenbecker P M, Haagen R, Lodomez S, Marsden P K and Schulz V 2011 A method for measuring the sub-pixel light distribution of scintillation detectors with digital SiPMs Peter. *IEEE Nuclear Science Symp. Conf. Record* pp 2301–2
- [80] Mora A D, Sieno L D, Re R, Pifferi A and Contini D 2020 Time-gated single-photon detection in time-domain diffuse optics: a review *Appl. Sci.* **10** 1101
- [81] Liu C and Berg R W 2012 Determining the spectral resolution of a charge-coupled device (CCD) Raman instrument *Appl. Spectrosc.* **66** 1034–43
- [82] Nissinen I, Lansman A K, Nissinen J, Holma J and Kostamovaara J 2013 2x(4x)128 time-gated CMOS single photon avalanche diode line detector with 100 ps resolution for Raman spectroscopy. *European Solid-State Circuits Conf.* pp 291–4

- [83] Maruyama Y, Blacksberg J and Charbon E 2014 A 1024× 8, 700-ps time-gated spad line sensor for planetary surface exploration with laser Raman spectroscopy and libs *IEEE J. Solid-State Circuits* **49** 179–89
- [84] Li Z and Deen M J 2014 Towards a portable Raman spectrometer using a concave grating and a time-gated CMOS SPAD *Opt. Express* **22** 18736
- [85] Heilala B, Mäkinen A, Nissinen I, Nissinen J, Mäkynen A and Perämäki P 2018 Evaluation of time-gated Raman spectroscopy for the determination of nitric, sulfuric and hydrofluoric acid concentrations in pickle liquor *Microchem. J.* **137** 342–7
- [86] Gyakwaa F, Aula M, Alatarvas T, Vuolio T, Huttula M and Fabritius T 2019 Applicability of time-gated Raman spectroscopy in the characterisation of calcium-aluminate inclusions *ISIJ Int.* **59** 1846–52
- [87] Wang Z, Shu Q and Chou K 2011 Structure of cao-B₂O₃-sio₂-tio₂ glasses: a Raman spectral study *ISIJ Int.* **51** 1021–7
- [88] Zheng K, Liao J, Wang X and Zhang Z 2013 Raman spectroscopic study of the structural properties of CaO-MgO-SiO₂-TiO₂ slags *J. Non. Cryst. Solids* **376** 209–15
- [89] Huang W J, Zhao Y H, Yu S, Zhang L X, Ye Z C, Wang N and Chen M 2016 Viscosity property and structure analysis of FeO-SiO₂-V₂O₃-TiO₂-Cr₂O₃ slags *ISIJ Int.* **56** 594–601
- [90] Shen X, Chen M, Wang N and Wang D 2019 Viscosity property and melt structure of CaO-MgO-SiO₂-Al₂O₃-FeO slag system *ISIJ Int.* **59** 9–15
- [91] Koivikko N, Laitinen T, Mouammine A, Ojala S and Keiski R L 2018 Catalytic activity studies of Vanadia/Silica-Titania catalysts in SVOC partial oxidation to formaldehyde: focus on the catalyst composition *Catalysts* **8** 56
- [92] El Assal Z, Ojala S, Zbair M, Echchtouki H, Nevanperä T, Pitkäaho S, Pirault-Roy L, Bensitel M, Brahmi R and Keiski R L 2019 Catalytic abatement of dichloromethane over transition metal oxide catalysts: thermodynamic modelling and experimental studies *J. Clean. Prod.* **228** 814–23
- [93] Tanskanen P, Heilala B, Kurki L, Lamberg P, Savela J and Fabritius T 2017 On-line monitoring of spodumene heat treatment process with time-gated Raman spectroscopy: *Process Mineralogy '17* (Cape Town)
- [94] Macleod N A and Matousek P 2008 Emerging non-invasive Raman methods in process control and forensic applications *Pharm. Res.* **25** 2205–15
- [95] Eliasson C, Claybourn M and Matousek P 2007 Deep subsurface Raman spectroscopy of turbid media by a defocused collection system *Appl. Spectrosc.* **61** 1123–7
- [96] Gal-Or E, Gershoni Y, Scotti G, Nilsson S M E, Saarinen J, Jokinen V, Strachan C J, Boije Af Gennäs G, Yli-Kauhaluoma J and Kotiaho T 2019 Chemical analysis using 3D printed glass microfluidics *Anal. Methods* **11** 1802–10
- [97] Dou J, Xu W, Koivisto J J, Mobley J K, Padmakshan D, Kögler M, Xu C, Willför S, Ralph J and Vuorinen T 2018 Characteristics of hot water extracts from the bark of cultivated willow (*Salix* sp.) *ACS Sustain. Chem. Eng.* **6** 5566–73
- [98] Yanagisawa S, Deshpande M S, Hirota S, Nakagawa T and Ogura T 2017 Improved stopped-flow time-resolved resonance Raman spectroscopy device for studying enzymatic reactions *J. Raman Spectrosc.* **48** 680–5
- [99] Virtanen T, Reinikainen S P, Kögler M, Mänttari M, Viitala T and Kallioinen M 2017 Real-time fouling monitoring with Raman spectroscopy *J. Memb. Sci.* **525** 312–9
- [100] Mesu J G, Visser T, Soulimani F and Weckhuysen B M 2005 Infrared and Raman spectroscopic study of pH-induced structural changes of L-histidine in aqueous environment *Vib. Spectrosc.* **39** 114–25
- [101] Kuball M, Riedel G J, Pomeroy J W, Sarua A, Uren M J, Martin T, Hilton K P, Maclean J O and Wallis D J 2007 Time-resolved temperature measurement of AlGaN/GaN electronic devices using micro-Raman spectroscopy *IEEE Electron Device Lett.* **28** 86–89
- [102] Sharma S K, Misra A K, Clegg S M, Barefield J E, Wiens R C and Acosta T 2010 Time-resolved remote Raman study of minerals under supercritical CO₂ and high temperatures relevant to Venus exploration *Philos. Trans. R. Soc. A Math. Phys. Eng. Sci.* **368** 3167–91
- [103] Fotso Gueutue E S, Canizares A, Simon P, Raimboux N, Hennem L and Ammar M R 2018 Nanosecond time-resolved Raman spectroscopy for solving some Raman problems such as luminescence or thermal emission *J. Raman Spectrosc.* **49** 822–9
- [104] Hooijschuur J H, Iping Petterson I E, Davies G R, Gooijer C and Ariese F 2013 Time resolved Raman spectroscopy for depth analysis of multi-layered mineral samples *J. Raman Spectrosc.* **44** 1540–7
- [105] Garcia C S, Abedin M N, Ismail S, Sharma S K, Misra A K, Nguyen T and Elsayed-Ali H 2009 Study of minerals, organic, and biogenic materials through time-resolved Raman spectroscopy *Proc. SPIE* (<https://doi.org/10.1117/12.818169>)
- [106] Romppanen S, Häkkinen H, Kekkonen J, Nissinen J, Nissinen I, Kostamovaara J and Kaski S 2019 Time-gated Raman and laser-induced breakdown spectroscopy in mapping of eudialyte and catapleite *J. Raman Spectrosc.* **50** 1462–9
- [107] Kekkonen J, Finnilä M A J, Heikkilä J, Anttonen V and Nissinen I 2019 Chemical imaging of human teeth by a time-resolved Raman spectrometer based on a CMOS single-photon avalanche diode line sensor *Analyst* **144** 6089–97
- [108] Usai A, Finlayson N, Gregory C D, Campbell C and Henderson R K 2019 Separating fluorescence from Raman spectra using a CMOS SPAD TCSPC line sensor for biomedical applications *Proc. SPIE* **10873** p 108730R
- [109] Knorr F, Smith Z J and Wachsmann-Hogiu S 2010 Development of a time-gated system for Raman spectroscopy of biological samples *Opt. Express* **18** 20049
- [110] Morris M D, Draper E R C, Goodship A E, Matousek P, Towrie M, Parker A W and Camacho N P 2005 Picosecond time-gated Raman spectroscopy for transcutaneous evaluation of bone composition *Proc. SPIE* pp 344
- [111] Kögler M, Paul A, Anane E, Birkholz M, Bunker A, Viitala T, Maiwald M, Junne S and Neubauer P 2018 Comparison of time-gated surface-enhanced Raman spectroscopy (TG-SERS) and classical SERS based monitoring of *Escherichia coli* cultivation samples *Biotechnol. Prog.* **34** 1533–42
- [112] Zeng F, Mou T, Zhang C, Huang X, Wang B, Ma X and Guo J 2019 Paper-based SERS analysis with smartphones as Raman spectral analyzers *Analyst* **144** 137–42
- [113] Mu T, Li S, Feng H, Zhang C, Wang B, Ma X, Guo J, Huang B and Zhu L 2019 High-sensitive smartphone-based Raman system based on cloud network architecture *IEEE J. Sel. Top. Quantum Electron.* **25** 1–7
- [114] Nelson A E 2020 The supercam instrument for the Mars 2020 Rover: *IEEE Aerospace Conf. 2020* vol 15 (Los Alamos, NM, United States) (Los Alamos, NM: Los Alamos National Lab. (LANL))
- [115] Uusitalo S *et al* 2019 Surface-enhanced Raman spectroscopy for beverage spoilage yeasts and bacteria detection with patterned substrates and gold nanoparticles (Conference Presentation) *Proc. SPIE* **10907** 1090703

- [116] Kögler M *et al* 2018 Bare laser-synthesized Au-based nanoparticles as non-disturbing SERS probes for Bacteria Identification *J. Organ. Behav.* **28** 303–25
- [117] Harz M, Rösch P and Popp J 2009 Vibrational spectroscopy-a powerful tool for the rapid identification of microbial cells at the single-cell level *Cytom. Part A* **75** 104–13
- [118] Jarvis R M and Goodacre R 2008 Characterisation and identification of bacteria using SERS. *Chem. Soc. Rev.* **37** 931–6
- [119] Buckley K and Ryder A G 2017 Applications of Raman spectroscopy in biopharmaceutical manufacturing: a short review *Appl. Spectrosc.* **71** 1085–116
- [120] Schlücker S 2014 Surface-enhanced Raman spectroscopy: concepts and chemical applications *Angew. Chemie., Int. Ed.* **53** 4756–95
- [121] Petry R, Schmitt M and Popp J 2003 Raman spectroscopy—a prospective tool in the life sciences. *Chemphyschem* **4** 14–30
- [122] Witkowska E *et al* 2017 Surface-enhanced Raman spectroscopy introduced into the International Standard Organization (ISO) regulations as an alternative method for detection and identification of pathogens in the food industry *Anal. Bioanal. Chem.* **409** 1555–67
- [123] Hobro A J and Lendl B 2009 Stand-off Raman spectroscopy *TrAC - Trends Anal. Chem.* **28** 1235–42
- [124] Gares K L, Hufziger K T, Bykov S V and Asher S A 2016 Review of explosive detection methodologies and the emergence of standoff deep UV resonance Raman *J. Raman Spectrosc.* **47** 124–41
- [125] Sadate S, Kassu A, Farley C W, Sharma A, Hardisty J and Lifson M T K 2011 Standoff Raman measurement of nitrates in water *Remote Sens. Model. Ecosyst. Sustain. VIII* **8156** 81560D
- [126] Hirschfeld T, Schildkraut E R, Tannenbaum H and Tanenbaum D 1973 Remote spectroscopic analysis of ppm-level air pollutants by Raman spectroscopy *Appl. Phys. Lett.* **22** 38–40
- [127] Skulinova M *et al* 2014 Time-resolved stand-off UV-Raman spectroscopy for planetary exploration *Planet. Space Sci.* **92** 88–100
- [128] Kekkonen J, Nissinen J, Kostamovaara J and Nissinen I 2018 Distance-resolving Raman radar based on a time-correlated CMOS single-photon avalanche diode line sensor *Sensors* **18** 1–11
- [129] Xia J, Yao Q, Zhu L, Dong M and Lou X 2019 Performance analysis and small signal identification of time-resolved stand-off Raman spectroscopy system *Vib. Spectrosc.* **102** 16–23
- [130] Gulati K K, Gulia S, Gambhir T, Kumar N, Gambhir V and Reedy M N 2019 Standoff detection and identification of explosives and hazardous chemicals in simulated real field scenario using time gated Raman spectroscopy *Def. Sci. J.* **69** 342–7